

**NPL REPORT
DQL-AS 030**

**Interlaboratory Comparison
of Reduced Modulus
Measurement of Polymers
at the Nanoscale Using an
AFM or a Nanoindenter**

- Protocol For Analysis

C A Clifford and M P Seah

July 2006



The DTI drives our ambition of 'prosperity for all' by working to create the best environment for business success in the UK. We help people and companies become more productive by promoting enterprise, innovation and creativity.

We champion UK business at home and abroad. We invest heavily in world-class science and technology. We protect the rights of working people and consumers. And we stand up for fair and open markets in the UK, Europe and the world.

The National Physical Laboratory is operated on behalf of the DTI by NPL Management Limited, a wholly owned subsidiary of Serco Group plc

**Interlaboratory Comparison of Reduced Modulus Measurement of
Polymers at the Nanoscale Using an AFM or a Nanoindenter
- Protocol For Analysis**

C A Clifford and M P Seah
Quality of Life Division
National Physical Laboratory
Teddington, Middlesex, TW11 0LW, UK
charles.clifford@npl.co.uk

ABSTRACT

This report describes the protocol for analysis for an interlaboratory comparison of reduced modulus measurement of polymers at the nanoscale using an AFM or a Nanoindenter. Five samples are provided for this study, along with this protocol, which contains simple instructions for the analysis.

© Crown copyright 2006
Reproduced with the permission of the Controller of HMSO
and Queen's Printer for Scotland

ISSN 1744-0602

National Physical Laboratory
Hampton Road, Teddington, Middlesex, TW11 0LW

Extracts from this report may be reproduced provided the source is acknowledged and the
extract is not taken out of context.

Approved on behalf of the Managing Director, NPL
by Dr I Gilmore, Knowledge Leader, Analytical Science Team

CONTENTS

1	OBJECTIVES	1
2	TIME TABLE	1
3	THIS PACKAGE	1
4	THE MATERIAL	1
5	SAMPLE HANDLING	2
6	PRACTICAL ISSUES IN MEASURING THE REDUCED MODULUS	2
7	REPORTING AND CALCULATING THE RESULTS	4
8	DATA REPORTING FORMAT	5
9	ANY DOUBTS	5
	ACKNOWLEDGEMENTS	5
	REFERENCES	5

1 OBJECTIVES

The measurand in this study is the reduced modulus in the surface or near surface region of each of a total of 5 samples in order to (i) to evaluate the current comparability of reduced modulus measurement at the nanoscale, (ii) to evaluate the repeatability of reduced modulus measurements.

2 TIME TABLE

You should complete the analysis for this work by 31 August 2006. If you cannot do so and need extra time please inform charles.clifford@npl.co.uk.

3 THIS PACKAGE

This package, unless you have otherwise specified, contains this protocol and a set of 5 Fluoroware containers, each containing one sample. Inspect the packaging to check if it has been opened by customs and if the integrity of the samples has been compromised. Please notify us that everything is received in order.

I have emailed NPL that all is OK with the samples on / / 2006

4 THE MATERIAL

The materials for analysis in the standard set comprise one piece of each of low-density polyethylene (LDPE), polypropylene (PP), polymethylpentene (PMP) polycarbonate (PC), polystyrene (PS). The 5 pieces of material are each supplied **face down** in separate containers to maintain their identity and integrity.

Your sample codes are given in the covering letter in a completed table of the form:

Serial Numbers of Your Samples

Polymer				
LDPE (0.2 GPa)	PP (1.4 GPa)	PMP (1.7 GPa)	PC (2.5 GPa)	PS (3.3 GPa)

Also included in the table are nominal values of the reduced modulus of each sample taken from the literature and given as a guide in case you need to pre-select instrumental options or parameters. Each sample, unless you have otherwise specified, is approximately 1 cm² in area. All measurements should take place on the smoother or shinier side of the samples.

Do not unpack these samples until you are nearly ready to analyse them. Before doing so, read the rest of this protocol so that you are clear about the next steps.

5 SAMPLE HANDLING

The following comments should prove helpful in handling samples.

If you intend to analyse the samples one by one by mounting them on the same sample holder, it is good practice to keep the samples in their containers and remove them only when you are ready to mount them. If you mount them together on one sample holder, ensure that their identities are maintained.

Inspect the samples for any scratches, blemishes or marks on the shinier surface. There may be some scratches on some of the samples. Choose as featureless an area as possible for the analysis. Note the condition of the surface.

You may have appropriate sample handling procedures as routine in your laboratory. Keep to those procedures but ensure that they are always superior to the minimum level given below.

You may need to handle the samples using uncoated clean stainless steel tweezers. If so, grip the sample at the edge only, in a region that will not be analysed. Avoid breathing or speaking over the samples. Use polythene gloves, or gloves of a higher quality, to avoid contaminating the tweezers or your cleaning equipment with finger grease. Do not use moulded gloves, for example vinyl, which may be covered with highly contaminating release agents.

6 PRACTICAL ISSUES IN MEASURING THE REDUCED MODULUS

If necessary, remove any dust or particulate matter on the surface [use, for instance, pure dry inert gases such as Argon or nitrogen to remove any particles. Do not use gas from pressurised cans that include a propellant or from compressed air lines that may contain oil vapour, do not use a method likely to contaminate the sample]. Do not use any other cleaning method as this may affect the results.

Ensure that any necessary instrument calibrations are up to date. For AFM this may include

z piezoelectric scanner calibration, tip shape, spring constant [1] and laser sensitivity. For nanoindenters this may include tip shape, force and displacement. For issues relating to calibration, please use the methods that you normally employ. If in doubt, some methods are given in reference [2-6].

To mount the samples, in order of preference the following methods are recommended: 1) use of a vacuum stage or a very secure mechanical holder, 2) epoxy glue (e.g. araldite), 3) cyanoacrylate glue (superglue). Sticky tape should be avoided if at all possible. Please note the holding mechanism in your report.

Conduct your analyses of the 5 samples. Seven measurements are required at two indentation depths for each sample and for two instrumental conditions, as described in the following sections.

Try to ensure that your analyses are in the central regions of the samples in a featureless area of the sample. Do not acquire data within 1 mm of the sample edge. Note the order of analysis of the samples. Record all of the instrumental parameters used. Record the temperature and humidity of the laboratory at the time of the experiment and whether this was similar or noticeably different to the temperature and humidity in the 2 days leading up to the experiment.

For the analysis, do 2 experiments under the following conditions

i) FIXED CONDITIONS FOR ALL 5 SAMPLES

Allow the instrument to warm up for at least one hour. For AFM do this with the laser aligned on the cantilever. Please choose a suitable tip shape and radius to use. For AFM, a stiff cantilever is required, some of the samples may be too stiff for indentation with lower spring constants and larger radius tips.

Undertake any necessary calibrations, e.g tip shape, laser sensitivity, etc.

For each sample, land on the sample with the lowest set-force needed to keep the tip in contact. Choose an area with minimal surface features. Undertake the indentation experiments with:

1) A constant load and unload rate with no hold time, so that the total load/unload cycle lasts 20 seconds (0.05 Hz).

2) Conduct measurements at two indentation depths of nominally 50 and 150 nm with 7 indents at each depth.

3) Keep the indents well spaced from each other. A pattern you may choose, depending on the radius of your tip is to: land at (0,0), and indent at the following positions in micrometres from the landing point, for the first indentation depth: (0, 4), (4, 4), (4, 0), (4, -4), (0, -4), (-4,-4), (-4, 0), and for the second indentation depth: (-4, 4), (-4, 8), (0, 8), (4, 8), (8, 8), (8, 4), (8, 0). If you wish to conduct more analyses continue with (8,-4), (8,-8), etc as shown in figure 1.

8 DATA REPORTING FORMAT

In decreasing order of preference, the preferred digital data formats are:

- (1) Three column ASCII format as shown in Table 1.
- (2) Manufacturer's text data format.
- (3) A simple binary format – with details of structure supplied.

Table 1 - Format for ASCII data

Column 1	Column 2	Column 2
Indentation depth (nm)	Force (nN)	Time (s)
0.113	0.000	0.000
0.325	14.881	0.014
0.654	29.763	0.029

Send your report and data preferably electronically or if not hardcopy to:

C A Clifford, National Physical Laboratory, Hampton Road, Teddington, Middlesex TW11 0LW, UK, email: charles.clifford@npl.co.uk
(Tel: +44 20 8943 6620; Fax: +44 20 8943 6453)

9 ANY DOUBTS

If you are unsure of anything or need advice please contact Charles Clifford, email: charles.clifford@npl.co.uk.

ACKNOWLEDGEMENTS

This work forms part of the Valid Analytical Measurement programme supported by the United Kingdom Department of Trade and Industry.

REFERENCES

- [1] Clifford C A and Seah M P 2005 *Nanotechnology* **16** 1666
- [2] Clifford C A and Seah M P 2005 *Appl Surf Sci* **252** 1915
- [3] Vanlandingham M R, Villarrubia J S, Guthrie W F and Meyers G F 2001 *Macromol Symp* **167** 15
- [4] Oliver W C and Pharr G M 1992 *J Mat Res* **7** 1564
- [5] Oliver W C and Pharr G M 2004 *J Mat Res* **19** 3
- [6] Fischer-Cripps A C 2000 *Vacuum* **58** 569